

MANI. Trocar Kit


☐ Trocar with valved cannula

Actual size









■MANI® Trocar is designed to be extracted smoothly from the valved cannula without forceps.

In case of any difficulty in detaching the cannula from the trocar after the incision, hold the valve gently with tips of forceps and pull the blade out from the sclera.

Do not hold the supportive arms at the tip of handle.

•	Trocar	blade in	the shap	pe of MV	R knife.
	I IItimot	to oborni		ntina nro	wan taah

- Ultimate sharpness adopting proven technologies of MANI® Ophthalmic Knife.
- · Valved cannula for the simpler operation.
- · Sterile individual pack (1 pc.) for additional requirement or backup.

Product Name	Packaging	Color Coding	Order #
Trocar 25G S	6 Pcs. / Box		MTR25S
Trocar 23G S	6 Pcs. / Box		MTR23S

Infusion Cannula





Connection with the valved cannula

Product Name	Packaging	Order #
Infusion Cannula 25G S	6 Pcs. / Box	MIC25S
Infusion Cannula 23G S	6 Pcs. / Box	MIC23S

- · Attaching directly to the valved cannula.
- · Sterile individual pack (1 pc.) for additional requirement

- Trocar with valved cannula & Infusion cannula



Product Name	Packaging	Order #
Trocar Kit 25G S	3 Kits / Box	MTK25S
Trocar Kit 23G S	3 Kits / Box	MTK23S

 Sterile 1 kit consists of 3 pcs. trocar with the valved cannula and 1 pc. infusion cannula.

STERILE EO Disposable 2

Do not use if package is damaged or opened.

⚠ Read instructions before use.

C E 0197

Distributed by



8-3 Kiyohara Industrial Park, Utsunomiya, Tochigi, 321-3231 JAPAN TEL: +81.28.667.7565 / FAX: +81.28.667.8305 URL: http://www.mani.co.jp/en/

E-mail:seg-ovs@ms.mani.co.jp S00730(161201)